

Mass Spectrometry

Analysis Information

Product Name: Maltotriose
Operator: David
Injection Date: 11/8/2021 10:23:34 AM
Inj. Volume: 1
Batch No.: 4


Acquisition Parameter

Probe: ESI
Gas Temp: 350°C
Gas Flow: 11L/min
Nebulizer: 50psi
VCap: 4000v
Fragmentor: 70v
Reference: m/z 121.050873, 922.009798
T.Flow: 0.3ml/min
B.conc: 20%H₂O/80%ACN

Result

211108 1109-28-0 18 (0.646) Cm (15:20)

1: Scan ES+
3.13e6



Molecular Weight: 504.44

m/z([M+Na]⁺) Base Peak: 527.2

Caution

FOR RESEARCH PURPOSES ONLY.

NOT FOR HUMAN, VETERINARY DIAGNOSTIC OR THERAPEUTIC USE.

APEX BIO Technology

www.apexbt.com

7505 Fannin street, Suite 410, Houston, TX 77054.

Tel: +1-832-696-8203 | Fax: +1-832-641-3177 | Email: info@apexbt.com

